

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130 178</u> ✓	
Name <u>James Kimura</u>		ID# <u>5677</u>	Date <u>12/19/08</u> ✓
A	Agency <u>Alaska State Troopers</u>		Phone # <u>225-5118</u>
Instrument Location <u>Ketchikan AST Post</u> ✓			
Alco S/N <u>X124670</u> ✓		Target Value <u>.075</u> ✓	High Pressure <u>600</u>
B	Alco Test Values		
	<u>.086</u> 1 st Alco	<u>.085</u> 2 nd Alco	
Signature <u>[Signature]</u> ✓		<u>[Signature]</u> 1/9/09	
(OVER)			


Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Kimura, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

1/10/09

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 178 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

DECEMBER 19, 2008 ✓ *COB 1/9/09*

OPERATOR'S NAME:
KIMURA/JAMES/K
OPERATOR'S NUMBER: 5677
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 000000
DEPT/AGENCY: KETO
CASE/REPORT: 00-000000
TEST TYPE: U
ALCO TARGET VALUE: .075 ✓
ALCO S/N: X124670

— BREATH ANALYSIS —

.075 ADJUSTED FOR 30.48 in
ALCO TARGET .076 21:39
BLANK TEST .000 21:40
INTERNAL STANDARD VERIFIED 21:40
ALCO TV 30.48 in .086 ✓ 21:41
BLANK TEST .000 *COB 1/9/09* 21:42
SUBJECT SAMPLE .000 21:42
BLANK TEST .000 21:43
ALCO TV 30.48 in .085 ✓ 21:43
BLANK TEST .000 21:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

DECEMBER 19, 2008 ✓
TIME 21:11

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.50 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢